







UK Standards for Microbiology Investigations

Catalase test



Acknowledgments

UK Standards for Microbiology Investigations (UK SMIs) are developed under the auspices of UKHSA working in partnership with the partner organisations whose logos are displayed below and listed on [the UK SMI website](#). UK SMIs are developed, reviewed and revised by various working groups which are overseen by a [steering committee](#).

The contributions of many individuals in clinical, specialist and reference laboratories who have provided information and comments during the development of this document are acknowledged. We are grateful to the medical editors for editing the medical content.

UK SMIs are produced in association with:



Displayed logos correct as of December 2024

Contents

Acknowledgments	2
Contents	3
Amendment table	4
1 General information	6
2 Scientific information	6
3 Scope of document	6
4 Introduction	6
5 Technical information/limitations	6
6 Safety considerations	7
7 Reagents and equipment	7
8 Quality control organisms	8
9 Procedure and results	8
Algorithm: Catalase test	10
References	11

Amendment table

Each UK SMI document has an individual record of amendments. The amendments are listed on this page. The amendment history is available from standards@ukhsa.gov.uk.

Any alterations to this document should be controlled in accordance with the local document control process.

Amendment number/date	8/18.02.25
Issue number discarded	4
Insert issue number	4.1
Section(s) involved	Amendment
Whole document.	<p>This is an administrative point change.</p> <p>The content of this UK SMI document has not changed.</p> <p>The last scientific and clinical review was conducted on 02.04.2019.</p> <p>Hyperlinks throughout document updated to Royal College of Pathologists website.</p> <p>Public Health England replaced with UK Health Security Agency throughout the document, including the updated Royal Coat of Arms</p> <p>Partner organisation logos updated.</p> <p>Broken links to devolved administrations replaced.</p> <p>References to NICE accreditation removed.</p> <p>Scope and Purpose replaced with General and Scientific information to align with current UK SMI template.</p>

Amendment number/date	7/02.04.19
Issue number discarded	3
Insert issue number	4
Anticipated next review date*	02.04.22
Section(s) involved	Amendment
Whole document.	Document and flowchart updated.

Catalase test

	<p>Technical limitations updated with subheadings.</p> <p>References updated with grades.</p> <p>Alternative positive bacterial NCTC strain (NCTC 12973) tested and validated for this test and EUCAST susceptibility tests.</p> <p>Fungal NCPF strains added.</p>
--	--

*Reviews can be extended up to 5 years where appropriate

1 General information

[View general information](#) related to UK SMIs.

2 Scientific information

[View scientific information](#) related to UK SMIs.


3 Scope of document

This test detects the catalase enzyme present in most cytochrome-containing aerobic and facultative anaerobic bacteria¹. *Streptococcus* and *Enterococcus* species are exceptions. Yeast such as *Cryptococcus neoformans* is catalase positive and can be presumptively identified using catalase test².

This UK SMI should be used in conjunction with other UK SMIs.

4 Introduction

The catalase test is used to detect the presence of catalase enzyme by the decomposition of hydrogen peroxide to release oxygen and water as shown by the following reaction:



The catalase reaction is evident by the rapid formation of bubbles.

Hydrogen peroxide is formed by some bacteria as an oxidative end product of the aerobic breakdown of sugars. If allowed to accumulate, it is highly toxic to bacteria and can result in cell death. Catalase either decomposes hydrogen peroxide or oxidises secondary substrates, but it has no effect on other peroxides².

There are method variations of the catalase test and these include the slide test method, the tube or bottle method and the agar slant method³. However, the commonly used methods in microbiology laboratories are the tube or bottle method and the agar slant method because it limits catalase aerosols, which have been shown to carry viable bacterial cells, that if inhaled could cause infections as well as contamination in other laboratory work being set up and work surface areas⁴.

5 Technical information/limitations

5.1 Interpretation of results

Media containing whole red blood cells will contain catalase and could therefore give a false positive result.

The enzyme, catalase is present in viable cultures only, so colony growth must be from an 18 to 24hr culture. Older cultures may lose their catalase activity and give false negative reactions².

Catalase test

Some inoculating loops or wires (nichrome) can react with the hydrogen peroxide to produce false positive reactions⁵.

False positive results can also be produced by dirty glass test tubes or bijoux bottles⁶.

5.2 False reactions

A weak catalase or pseudocatalase reaction may be produced by some strains of *Aerococcus* species. Some strains of *Enterococcus* species also produce a pseudocatalase.

Cultures of anaerobic bacteria should be exposed to air for 30 min prior to testing².

5.3 Quality control

Hydrogen peroxide is unstable and must be refrigerated at all times. Avoid any undue exposure to light.

6 Safety considerations⁷⁻²⁴

Refer to current guidance on the safe handling of all organisms and reagents documented in this UK SMI.

Catalase testing of bacteria can be hazardous due to the release of bacteria-laden aerosols by liberated oxygen⁴. All work likely to generate aerosols must be performed in a microbiological safety cabinet.

Hydrogen peroxide is a highly corrosive chemical (depending on the concentration); therefore, appropriate personal protective clothing must be worn at all times when in use. Extreme care must be taken by persons using this reagent.

The above guidance should be supplemented with local COSHH and risk assessments.

Compliance with postal and transport regulations is essential.

7 Reagents and equipment²

Discrete bacterial/yeast colonies on solid medium

Note: The catalase test should not be performed on colonies taken from media containing whole red blood cells because they contain catalase and could therefore give a false positive result. Colonies taken from chocolate agar plate may be tested as the blood cells have been destroyed².

Inoculated pure agar slant culture

Hydrogen peroxide solution, 3–6 %. Commercial preparations are available.

Clean capped test tubes (plastic or glass) or Bijoux bottles

Bacteriological straight platinum wire/loop or disposable alternative

8 Quality control organisms

Bacteria

Positive control:

Staphylococcus aureus NCTC 6571 or NCTC 12973

Negative control:

Streptococcus mitis NCTC 10712

Fungi

Positive control:

Cryptococcus neoformans NCPF 3168

Negative control:

Candida albicans NCPF 3281

Note 1: Hydrogen peroxide is unstable and so should undergo a quality control check daily or immediately prior to use. The positive and negative controls should be run simultaneously.

Note 2: These bacterial strains have been validated by NCTC to give this result.

Note 3: The fungal strains have not been validated by NCTC to give this result at the time of publication.

9 Procedure and results

9.1 Tube or bottle method³

- Place 4 to 5 drops of hydrogen peroxide solution in a test tube or bijoux bottle
- Carefully pick a colony to be tested with a wire/loop or disposable alternative
- Rub the colony on the inside wall of the bottle just above the surface of the hydrogen peroxide solution
- Cap the tube or bottle and tilt it to allow the hydrogen peroxide solution to cover the colony
- Observe for immediate bubble formation (effervescence)

9.2 Agar slant method^{2,6}

- Add 1.0mL of H₂O₂ directly onto an 18 to 24hr heavily inoculated pure culture grown on a nutrient agar slant and replace the cap
- Observe for immediate bubbling (effervescence)

For both methods,

Positive result

Vigorous bubbling indicates the presence of catalase.


Catalase test

Negative result

No bubbling indicates the absence of catalase.

Note: Both positive and negative controls must be tested alongside the test organism.

Algorithm: Catalase test



Note:

Bacteria

Positive Control: *Staphylococcus aureus* NCTC 6571 or NCTC 12973

Negative Control: *Streptococcus mitis* NCTC 10712

Fungi

Positive control: *Cryptococcus neoformans* NCPF 3168

Negative control: *Candida albicans* NCPF 3281

References

An explanation of the reference assessment used is available in the [scientific information section on the UK SMI website](#).

1. Doelle HW. Chemosynthesis - aerobic respiration. In: Doelle HW, editor. Bacterial Metabolism. London: Academic Press; 1969. **B, IV**
2. MacFaddin JF. Catalase-Peroxidase Tests. Biochemical Tests for Identification of Medical Bacteria. 3rd ed. Philadelphia: Lippincott Williams and Wilkins; 2000. p. 78-97. **B, III**
3. Reiner K. Catalase Test Protocol. ASM Conference for Undergraduate Educators 2010. 2012. **B, VIII**
4. Duke PB, Jarvis JD. The catalase test--a cautionary tale. Med Lab Technol 1972;29:203-4. **B, III**
5. Ochei J, Kolhatkar A. Identification Methods. Medical Laboratory Science Theory and Practice; 2000. p. 644-58. **B, III**
6. Barrow GI, Feltham RKA Cowan and Steel's Manual for the Identification of Medical bacteria. 3rd ed.: Cambridge University Press; 2003. 78-97. **B, III**
7. Advisory Committee on Dangerous Pathogens. Infections at work: Controlling the risks. Her Majesty's Stationery Office 2003. **A, VI**
8. Advisory Committee on Dangerous Pathogens. Biological agents: Managing the risks in laboratories and healthcare premises. Health and Safety Executive 2005. **A, VI**
9. Advisory Committee on Dangerous Pathogens. Biological Agents: Managing the Risks in Laboratories and Healthcare Premises. Appendix 1.2 Transport of Infectious Substances - Revision. Health and Safety Executive 2008. **A, VI**
10. Advisory Committee on Dangerous Pathogens. The Approved List of Biological Agents. Health and Safety Executive 2013. 1-35. **A, VI**
11. British Standards Institution (BSI). BS EN12469 - Biotechnology - performance criteria for microbiological safety cabinets 2000. **A, VI**
12. British Standards Institution (BSI). BS 5726:2005 - Microbiological safety cabinets. Information to be supplied by the purchaser and to the vendor and to the installer, and siting and use of cabinets. Recommendations and guidance. 2005. 1-14. **A, VI**
13. Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. MMWR Surveill Summ 2012;61:1-102. **B, V**

14. Department for Transport. Transport of Infectious Substances, 2011 Revision 5. 2011. **A, VI**
15. Department of Health. Transport of Infectious Substances. Best Practice Guidance for Microbiology Laboratories. Department of Health. 1-13. 2007. **A, VI**
16. European Parliament. UK Standards for Microbiology Investigations (UK SMIs) use the term "CE marked leak proof container" to describe containers bearing the CE marking used for the collection and transport of clinical specimens. The requirements for specimen containers are given in the EU *in vitro* Diagnostic Medical Devices Directive (98/79/EC Annex 1 B 2.1) which states: "The design must allow easy handling and, where necessary, reduce as far as possible contamination of, and leakage from, the device during use and, in the case of specimen receptacles, the risk of contamination of the specimen. The manufacturing processes must be appropriate for these purposes". 1998. **A, VI**
17. Health and Safety Executive. Five Steps to Risk Assessment: A Step by Step Guide to a Safer and Healthier Workplace. HSE Books,. 2002. **A, VI**
18. Health and Safety Executive. A Guide to Risk Assessment Requirements: Common Provisions in Health and Safety Law. HSE Books,. 2002. **A, VI**
19. Health and Safety Executive. Safe use of pneumatic air tube transport systems for pathology specimens. 2009. **A, VI**
20. Health and Safety Executive. Control of Substances Hazardous to Health Regulations. The Control of Substances Hazardous to Health Regulations 2002 (as amended). Approved Code of Practice and guidance L5 (sixth edition). HSE Books,. 2013. **A, VI**
21. Health Services Advisory Committee. Safe Working and the Prevention of Infection in Clinical Laboratories and Similar Facilities. HSE Books 2003. **A, VI**
22. Home Office. Anti-terrorism, Crime and Security Act. 2001. **A, VI**
23. Official Journal of the European Communities. Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on *in vitro* diagnostic medical devices 1998. 1-37. **A, VI**
24. World Health Organization. Guidance on regulations for the Transport of Infectious Substances 2017-2018. 2017. **A, VI**