








UK Standards for Microbiology Investigations

Indole test



Acknowledgments

UK Standards for Microbiology Investigations (UK SMIs) are developed under the auspices of UKHSA working in partnership with the partner organisations whose logos are displayed below and listed on [the UK SMI website](#). UK SMIs are developed, reviewed and revised by various working groups which are overseen by a [steering committee](#).

The contributions of many individuals in clinical, specialist and reference laboratories who have provided information and comments during the development of this document are acknowledged. We are grateful to the medical editors for editing the medical content.

UK SMIs are produced in association with:



Displayed logos correct as of December 2024

Contents

Acknowledgments	2
Contents	3
Amendment table	4
1 General information	6
2 Scientific information	6
3 Scope of document	6
4 Introduction	6
5 Technical information/limitations	6
6 Safety considerations	7
7 Reagents and equipment	8
8 Quality control organisms	8
9 Procedure and results	8
Algorithm: Indole test	10
References	11

Amendment table

Each UK SMI document has an individual record of amendments. The amendments are listed on this page. The amendment history is available from standards@ukhsa.gov.uk.

Any alterations to this document should be controlled in accordance with the local document control process.

Amendment number/date	8/28.02.25
Issue number discarded	4
Insert issue number	4.1
Section(s) involved	Amendment
Whole document.	<p>This is an administrative point change.</p> <p>The content of this UK SMI document has not changed.</p> <p>The last scientific and clinical review was conducted on 03/12/2018.</p> <p>Hyperlinks throughout document updated to Royal College of Pathologists website.</p> <p>Public Health England replaced with UK Health Security Agency throughout the document, including the updated Royal Coat of Arms</p> <p>Partner organisation logos updated.</p> <p>Broken links to devolved administrations replaced.</p> <p>References to NICE accreditation removed.</p> <p>Scope and Purpose replaced with General and Scientific information to align with current UK SMI template.</p>

Amendment number/date	7/03.12.18
Issue number discarded	3
Insert issue number	4
Anticipated next review date*	03.12.21
Section(s) involved	Amendment
Whole document.	Document and flowchart updated.

Indole test

	Technical limitations updated with subheadings. References updated with grades.
Quality control organisms.	Alternative positive bacterial NCTC strain tested and validated for this test.

*Reviews can be extended up to 5 years where appropriate

1 General information

[View general information](#) related to UK SMIs.

2 Scientific information

[View scientific information](#) related to UK SMIs.


3 Scope of document

This document covers the procedure for indole test. The indole test detects tryptophanase production and is an aid in the differentiation of the Enterobacterales and other genera.

This UK SMI should be used in conjunction with other UK SMIs.

4 Introduction

The indole test determines the ability of an organism to produce indole from the degradation of the amino acid tryptophan. Tryptophan is hydrolysed by tryptophanase to produce three possible end products – one of which is indole, the others are pyruvate and ammonium ion as shown by the following reaction¹:



A coloured product is produced when the indole is combined with certain aldehydes².

Two indole test methods are described; a spot indole test, which detects rapid indole producing organisms and a conventional tube method requiring overnight incubation, which identifies weak indole producing organisms.

5 Technical information/limitations

5.1 Peptone broth varieties

If peptone broth is used instead of tryptophan broth, the batch should be checked with a positive control to ensure the peptone is adequate for indole production. This is because there are varieties of peptone broth media on the market, and some are unsuitable for indole production because they contain too little tryptophan.

5.2 Spot indole method

Organisms to be tested by the spot indole method must be taken from a tryptophan - containing medium (for example blood agar) and never from MacConkey agar as they have pH indicators and pigmentation of lactose-positive colonies which will make interpretation of colour reaction difficult¹. The test can be carried out from some chromogenic agars^{3,4}.

Indole is a diffusible product. To mitigate indole diffusion, select a well isolated colony for the spot indole test.

5.3 Inhibition of indole production

Peptone media with added glucose should not be used because acid production may inhibit indole production due to a change in pH^{1,5}.

5.4 False reactions

Anaerobes, particularly *Clostridium* species, form indole but can rapidly break it down as it is produced; therefore, false negative reactions may occur¹.

False positive reactions may occur with the spot indole test if the inoculum is a mixed culture of indole positive and indole negative organisms^{1,4,6}.

5.5 Aerobic incubation

Cultures to be tested for indole must be incubated aerobically because a decrease in oxygen tension decreases indole production¹.

5.6 Alternative reagent

Ehrlich's reagent, an alternative to Kovács reagent, also contains Dimethylamino-benzaldehyde (DMAB), which reacts with indole to produce a red product. The Ehrlich formulation is more sensitive but contains additional toxic or flammable solvents; it is recommended when testing bacterial groups that produce little indole such as non-fermentative bacilli or anaerobes. Kovács reagent is more stable and the absence of the additional organic extraction (required with Ehrlich's) makes Kovács formulation more suitable for laboratories⁷.

6 Safety considerations⁸⁻²⁵

Refer to current guidance on the safe handling of all organisms and reagents documented in this UK SMI.

All work likely to generate aerosols must be performed in a microbiological safety cabinet.

Extreme care should be taken by staff when the Kovács's reagent has to be made up before use, as one of the key ingredients used is the concentrated Hydrochloric acid and it is highly corrosive.

Kovács's indole reagent is an irritant.

The above guidance should be supplemented with local COSHH and risk assessments.

Compliance with postal and transport regulations is essential.

7 Reagents and equipment

Discrete bacterial colonies on solid medium.

Tube method

1% tryptophan or peptone broth.

Kováč's reagent (for use with broth cultures).

Bacteriological straight wire/loop (preferably nichrome) or disposable alternative.

Spot indole test

Whatman no. 1 Filter paper.

Spot indole reagent (1% or 5% *p*-methylaminobenzaldehyde OR 1% *p*-dimethylaminocinnamaledehyde)⁴.

If using commercial kit, follow manufacturer's instructions.

Bacteriological straight wire/loop (preferably nichrome) or disposable alternative.

Petri dish.

8 Quality control organisms

Positive control:

Escherichia coli NCTC 10418 or NCTC 12241

Negative control:

Proteus mirabilis NCTC 10975

Note: The reference strains are validated by NCTC for the test shown.

9 Procedure and results

9.1 Tube method (broth cultures)^{1,26}

- inoculate the tryptophan (or peptone) broth with the test organism and incubate at 37°C for 24 - 48hr
- add 0.5mL of the Kováč's reagent and shake gently
- examine the upper layer of liquid after about 1min

Positive result

Formation of a pink to red colour (occurring within a few seconds)

Negative result

No colour change, the reagent layer remains yellow or slightly cloudy

9.2 Spot indole test^{4,27}

- place a piece of filter paper (Whatman no.1) into a sterile Petri dish and moisten with 1 -1.5mL Indole reagent or if using commercial pre-prepared filter paper containing the indole reagent, to equilibrate to room temperature before use
- smear an isolated pure colony (from an 18 -24hr culture) onto the saturated surface of the filter paper using a sterile loop
- examine immediately

Positive result

Follow manufacturer's instructions and interpretations.


Negative result

Follow manufacturer's instructions and interpretations.

Note:

1. The API commercial kits can also be used to determine whether an organism is Indole positive or negative.
2. Depending on the spot indole reagent used for the spot indole test, the resulting colours differ. If using *p*-methylaminobenzaldehyde, the presence of indole is indicated by a red colour and if using *p*-dimethylaminocinnamaledhyde, a bluish-green colour is observed.

Algorithm: Indole test



Note:

Positive control: *Escherichia coli* NCTC 10418 or NCTC 12241
Negative control: *Proteus mirabilis* NCTC 10975

References

An explanation of the reference assessment used is available in the [scientific information section on the UK SMI website](#).

1. MacFaddin JF. Indole Test. *Biochemical Tests for Identification of Medical Bacteria*. 3rd ed. Philadelphia: Lippincott Williams and Wilkins; 2000. p. 221-32. **B, III**
2. Bailey and Scott's. *Diagnostic Microbiology*. In: Forbes BA, Sahm DF, Weissfeld AS, editors. 11th ed. St Louis: Mosby Inc; 2002. p. 152-3. **B, III**
3. Perry JD, Butterworth LA, Nicholson A, Appleby MR, Orr KE. Evaluation of a new chromogenic medium, Uriselect 4, for the isolation and identification of urinary tract pathogens. *JClinPathol* 2003;56:528-31. **B, III**
4. Peterson WC, Hale DC, Matsen JM. An evaluation of the practicality of the spot-indole test for the identification of *Escherichia coli* in the clinical microbiology laboratory. *Am J Clin Pathol* 1982;78:755-8. **B, III**
5. Epps HM, Gale EF. The influence of the presence of glucose during growth on the enzymic activities of *Escherichia coli*: comparison of the effect with that produced by fermentation acids. *BiochemJ* 1942;36:619-23. **B, II**
6. Bale MJ, McLaws SM, Matsen JM. The spot indole test for identification of swarming *Proteus*. *Am J Clin Pathol* 1985;83:87-90. **B, III**
7. MacWilliams M. Indole Test Protocol. American Society for Microbiology Peer-reviewed. 2013. **B, VIII**
8. Advisory Committee on Dangerous Pathogens. *Infections at work: Controlling the risks*. Her Majesty's Stationery Office 2003. **A, VI**
9. Advisory Committee on Dangerous Pathogens. *Biological agents: Managing the risks in laboratories and healthcare premises*. Health and Safety Executive 2005. **A, VI**
10. Advisory Committee on Dangerous Pathogens. *Biological Agents: Managing the Risks in Laboratories and Healthcare Premises. Appendix 1.2 Transport of Infectious Substances - Revision*. Health and Safety Executive 2008. **A, VI**
11. Advisory Committee on Dangerous Pathogens. *The Approved List of Biological Agents*. Health and Safety Executive 2013. 1-35. **A, VI**
12. British Standards Institution (BSI). *BS EN12469 - Biotechnology - performance criteria for microbiological safety cabinets* 2000. **A, VI**
13. British Standards Institution (BSI). *BS 5726:2005 - Microbiological safety cabinets. Information to be supplied by the purchaser and to the vendor and to*

- the installer, and siting and use of cabinets. Recommendations and guidance. 2005. 1-14. **A, VI**
14. Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. MMWR Surveill Summ 2012;61:1-102. **B, V**
 15. Department for Transport. Transport of Infectious Substances, 2011 Revision 5. 2011. **A, VI**
 16. Department of Health. Transport of Infectious Substances. Best Practice Guidance for Microbiology Laboratories. Department of Health. 1-13. 2007. **A, VI**
 17. European Parliament. UK Standards for Microbiology Investigations (UK SMIs) use the term "CE marked leak proof container" to describe containers bearing the CE marking used for the collection and transport of clinical specimens. The requirements for specimen containers are given in the EU *in vitro* Diagnostic Medical Devices Directive (98/79/EC Annex 1 B 2.1) which states: "The design must allow easy handling and, where necessary, reduce as far as possible contamination of, and leakage from, the device during use and, in the case of specimen receptacles, the risk of contamination of the specimen. The manufacturing processes must be appropriate for these purposes". 1998. **A, VI**
 18. Health and Safety Executive. Five Steps to Risk Assessment: A Step by Step Guide to a Safer and Healthier Workplace. HSE Books,. 2002. **A, VI**
 19. Health and Safety Executive. A Guide to Risk Assessment Requirements: Common Provisions in Health and Safety Law. HSE Books,. 2002. **A, VI**
 20. Health and Safety Executive. Safe use of pneumatic air tube transport systems for pathology specimens. 2009. **A, VI**
 21. Health and Safety Executive. Control of Substances Hazardous to Health Regulations. The Control of Substances Hazardous to Health Regulations 2002 (as amended). Approved Code of Practice and guidance L5 (sixth edition). HSE Books,. 2013. **A, VI**
 22. Health Services Advisory Committee. Safe Working and the Prevention of Infection in Clinical Laboratories and Similar Facilities. HSE Books 2003. **A, VI**
 23. Home Office. Anti-terrorism, Crime and Security Act. 2001. **A, VI**
 24. Official Journal of the European Communities. Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on *in vitro* diagnostic medical devices 1998. 1-37. **A, VI**
 25. World Health Organization. Guidance on regulations for the Transport of Infectious Substances 2017-2018. 2017. **A, VI**

Indole test

26. Feltham RKA, Barrow GI. Cowan and Steel's Manual for the Identification of Medical Bacteria. 3rd ed. Cambridge: Cambridge University Press; 2003. p. 219-30- 31. **B, III**
27. Miller JM, Wright JW. Spot indole test: evaluation of four reagents. JClinMicrobiol 1982;15:589-92. **B, III**